



201719002006

检测报告

TEST REPORT

委托方 Commissioned by	广东爱彼士医疗器械科技有限公司 Guangdong aibishi medical equipment technology co., ltd
生产单位 Factory	广东爱彼士医疗器械科技有限公司 Guangdong aibishi medical equipment technology co., ltd
样品名称 Name of Sample	普通防护口罩 Disposible protective mask
型号规格 Type, Specification	大号
检测类别 Testing Purpose	委托检测 Commission



广东产品质量监督检验研究院

GUANGDONG TESTING INSTITUTE OF PRODUCT QUALITY SUPERVISION

检 测 报 告

TEST REPORT



序号 No	检测项目[单位] Test items[Unit]	标准条款 Standard terms	标准要求 Standard requirements	检测结果 Result	单项结论 Conclusion	备注 Notes
1	外观要求 Appearance requirements	5.2	口罩表面不得有破损、油污斑渍、变形及其他明显的缺陷 The surface of the mask should be free from damage, oil stains, deformation and other obvious defects	符合 Satisfactory	合格 Pass	/
2	颗粒过滤效率 [%] Particle Filtration Efficiency (%)	5.3	≥ 80	90.4	合格 Pass	/
				88.5		
				91.3		
3	通气阻力[Pa] Ventilation resistance(Pa)	5.3	≤ 80	54.3	合格 Pass	/
				59.2		
				56.1		
4	口罩带及口罩带与口罩体的连接处断裂强力[N] Breaking strength of mask string and junction between the mask string and mask body (N)	5.3	≥ 10	符合 Satisfactory	合格 Pass	/
				符合 Satisfactory		
				符合 Satisfactory		
5	口罩下方视野 Visual field	5.3	$\geq 60^\circ$	67°	合格 Pass	/
6	细菌过滤效率 (BFE) [%] Bacterial filtration efficiency (%)	5.3	$\geq 95\%$	99.9	合格 Pass	/
				99.9		
				99.9		

A certification is included in the box

合格证 CERTIFICATION
一次性民用防护口罩
DISPOSABLE PROTECTIVE MASK

规格:	170*95mm, 50pcs\box
Specification	
执行标准:	GB/T32610-2016
Executive Standard	
检验:	合格
Test Qualified	Qualified
原材料:	71%无纺布+29%熔喷布
Raw materials	71% non-woven fabric + 29% Melt-blown fabric
生产批号/Production batch:	20200422
生产日期/Production Date:	2020-4-22
有效期/Period of Validity:	2 years
生产企业:	广东爱彼士医疗器械科技有限公司
Manufacturer	GUANGDONG EBIS MEDICAL EQUIPMENT TECHNOLOGY CO., LTD. 中国广东省汕头市潮阳区和平镇新龙北新路工 业区523号
生产地址:	NO 523, INDUSTRIAL ZONE, XINLONGBEIXIN ROAD, HEPING TOWN, CHAOYANG DISTRICT, SHANGTOU CITY
Address	

本产品经检验符合规定的质量标准要求，准予出厂

Made in China